ACR-SPR-SSR PRACTICE PARAMETER FOR THE PERFORMANCE AND INTERPRETATION OF MAGNETIC RESONANCE IMAGING (MRI) OF THE ANKLE AND HINDFOOT

The American College of Radiology, with more than 40,000 members, is the principal organization of radiologists, radiation oncologists, and clinical medical physicists in the United States. The College is a nonprofit professional society whose primary purposes are to advance the science of radiology, improve radiologic services to the patient, study the socioeconomic aspects of the practice of radiology, and encourage continuing education for radiologists, radiation oncologists, medical physicists, and persons practicing in allied professional fields.

The American College of Radiology will periodically define new practice parameters and technical standards for radiologic practice to help advance the science of radiology and to improve the quality of service to patients throughout the United States. Existing practice parameters and technical standards will be reviewed for revision or renewal, as appropriate, on their fifth anniversary or sooner, if indicated.

Each practice parameter and technical standard, representing a policy statement by the College, has undergone a thorough consensus process in which it has been subjected to extensive review and approval. The practice parameters and technical standards recognize that the safe and effective use of diagnostic and therapeutic radiology requires specific training, skills, and techniques, as described in each document. Reproduction or modification of the published practice parameter and technical standard by those entities not providing these services is not authorized.

PREAMBLE

This document is an educational tool designed to assist practitioners in providing appropriate radiologic care for patients. Practice Parameters and Technical Standards are not inflexible rules or requirements of practice and are not intended, nor should they be used, to establish a legal standard of care 1. For these reasons and those set forth below, the American College of Radiology and our collaborating medical specialty societies caution against the use of these documents in litigation in which the clinical decisions of a practitioner are called into question. The ultimate judgment regarding the propriety of any specific procedure or course of action must be made by the practitioner considering all the circumstances presented. Thus, an approach that differs from the guidance in this document, standing alone, does not necessarily imply that the approach was below the standard of care. To the contrary, a conscientious practitioner may responsibly adopt a course of action different from that set forth in this document when, in the reasonable judgment of the practitioner, such course of action is indicated by variables such as the condition of the patient, limitations of available resources, or advances in knowledge or technology after publication of this document. However, a practitioner who employs an approach substantially different from the guidance in this document may consider documenting in the patient record information sufficient to explain the approach taken.

The practice of medicine involves the science, and the art of dealing with the prevention, diagnosis, alleviation, and treatment of disease. The variety and complexity of human conditions make it impossible to always reach the most appropriate diagnosis or to predict with certainty a particular response to treatment. Therefore, it should be recognized that adherence to the guidance in this document will not assure an accurate diagnosis or a successful outcome. All that should be expected is that the practitioner will follow a reasonable course of action based on current knowledge, available resources, and the needs of the patient to deliver effective and safe medical care. The purpose of this document is to assist practitioners in achieving this objective.

1 lowa Medical Society and Iowa Society of Anesthesiologists v. Iowa Board of Nursing, 831 N.W.2d 826 (Iowa 2013) Iowa Supreme Court refuses to find that the "ACR Technical Standard for Management of the Use of Radiation in Fluoroscopic Procedures (Revised 2008)" sets a national standard for who may perform fluoroscopic procedures in light of the standard's stated purpose that ACR standards are educational tools and not intended to establish a legal standard of care. See also, Stanley v. McCarver, 63 P.3d 1076 (Ariz. App. 2003) where in a concurring opinion the Court stated that "published standards or guidelines of specialty medical organizations are useful in determining the duty owed or the standard of care applicable in a given situation" even though ACR standards themselves do not establish the standard of care.

I. INTRODUCTION

This practice parameter was developed and written collaboratively by the American College of Radiology (ACR), the Society for Pediatric Radiology (SPR), and the Society of Skeletal Radiology (SSR).

Magnetic resonance imaging (MRI) is a proven imaging modality for the detection, evaluation, staging, and follow-up of disorders of the ankle and hindfoot. Properly performed and interpreted, MRI not only contributes

to diagnosis but also can guide treatment planning, help predict outcome, and increase diagnostic confidence [1-7]. However, ankle MRI should be performed only when clinically appropriate [8] and when the examination adds information not rendered from clinical examinations and radiographs. The strengths of MRI and other modalities should be weighed as to their suitability in particular patients and in particular clinical conditions.

Radiographs should be the first imaging test performed for suspected bone and soft-tissue abnormalities in the ankle and will often permit diagnosis or exclusion of an abnormality or will direct further imaging workup. Bone scintigraphy is most often used to screen the entire skeleton for multifocal diseases such as metastases, local conditions like complex regional pain syndrome, and other potentially radiographically occult bone disorders. Although MRI is typically the preferred imaging modality for suspected stress fractures and osteomyelitis of the foot and ankle, radionuclide imaging can also be useful to confirm or exclude these diagnoses [9]. Arthrography can also be used in a therapeutic role when combined with anesthetic and/or corticosteroid injection in localizing the source of pain [10-12]. Ultrasonography has come to play an increasingly important role in the diagnostic evaluation of the soft tissues of the ankle and foot, including tendons, ligaments, and soft-tissue masses [14-16]. Contrast tenography, which has been described in the evaluation and treatment of tenosynovitis in the hindfoot [17], has largely been replaced by sonography in centers in which this modality is performed. Stress radiography has been used with variable success in ankles with ligament injuries [18].

Computed tomography (CT), especially with multichannel row scanners and the use of multiplanar reformations, has an important role in the evaluation of complex fractures and dislocations of the ankle and hindfoot, osteochondral lesions, tarsal coalition, osteoid osteoma and for surgical planning [19-24]. Additionally, CT can be used for diagnosis of bone and soft-tissue injuries when there is a contraindication to MRI [25,26]. 3-D volume rendering can show the anatomic relationships of bones and tendons, which may be useful for preoperative planning [27,28]. When combined with arthrography, CT can also be used for evaluating the articular cartilage and joint bodies [29].

Although MRI is a sensitive, noninvasive diagnostic test for detecting anatomic abnormalities of the ankle and hindfoot, its findings may be misleading if not closely correlated with radiographs, clinical history, physical examination, physiologic tests such as nerve conduction analysis and electromyography, and other imaging studies when indicated. Adherence to the following parameters will increase the probability of detecting clinically important abnormalities.

II. INDICATIONS

- A. Primary indications for MRI of the ankle and hindfoot include, but are not limited to, diagnosis, exclusion, and grading of the following suspected disorders:
 - 1. Achilles tendon disorders: partial and complete tears, tendinitis, tendinopathy, treated tears, paratenonitis, and xanthomas [5,30-34]
 - 2. Posterior tibial tendon disorders: partial and complete tears, tendinitis, tendinopathy, tenosynovitis, subluxation, and dislocation [3,35-40]
 - 3. Peroneal tendon disorders: partial and complete tears, tendinitis, tendinopathy, tenosynovitis, subluxation, dislocation, and abnormalities of the peroneal retinaculum [14,41-44]
 - 4. Abnormalities of other hindfoot tendons: partial and complete tears, tendinitis, tendinopathy, tenosynovitis, and entrapment [11,45-49]
 - 5. Anterior and posterior talofibular, anterior and posterior tibiofibular, calcaneofibular, deltoid, spring, and syndesmotic ligament tears [6,13,18,50-58]
 - 6. Impingement syndromes: anterolateral, anteromedial, posterior, and posteromedial^{2,3} [13,40,59-67]
 - 7. Osteochondral abnormalities, degenerative or traumatic articular cartilage abnormalities, and intraarticular bodies³ [13,29,68-74]
 - 8. Neurologic conditions: nerve entrapment and compression, denervation neuropathy, including tarsal tunnel syndrome² [75-80]
 - 9. Plantar fasciitis, plantar fascia rupture, and plantar fibromatosis [81-84]
 - 10. Sinus tarsi syndrome² [85]
 - 11. Synovial-based disorders: inflammatory and nodular synovitis, tenosynovitis, bursitis, and ganglion cysts² [46,86-89]
 - 12. Marrow abnormalities: fractures, bone contusions, osteonecrosis, marrow edema syndromes, and stress fractures² [90-94]

- 13. Neoplasms of bone, joint, or soft tissue² [95-99]
- 14. Infections of bone, joint, or soft tissue² [100-103]
- 15. Congenital and developmental conditions: alignment anomalies and dysplasia, tarsal coalition, and symptomatic and asymptomatic normal variants such as accessory ossicles [76,104-109]
- B. MRI of the ankle and hindfoot may be indicated to further clarify and stage conditions diagnosed clinically and/or suggested by other imaging modalities, including, but not limited to:
 - 1. Arthritides: inflammatory, infectious, neuropathic, degenerative, crystal-induced, and posttraumatic² [4,36,46,87,110-114]
 - 2. Primary and secondary bone and soft-tissue tumors² [95-98,115] (see also the <u>ACR-SPR-SSR Practice Parameter for the Performance and Interpretation of Magnetic Resonance Imaging (MRI) of Bone and Soft-Tissue Tumors [116])</u>
 - 3. Fractures and stress fractures [117-119]
- C. MRI of the ankle and hindfoot may be useful to evaluate specific clinical scenarios, including, but not limited to:
 - 1. Prolonged, refractory, or unexplained ankle or heel pain^{2,3}
 - 2. Acute ankle trauma [7,55,120,121]
 - 3. Ankle and hindfoot injuries in athletes³ [90,122-125]
 - 4. Ankle or subtalar instability³ [6,51,55,126,127]
 - 5. Ankle and/or hindfoot malalignments [40,108]
 - 6. Limited or painful range of motion
 - 7. Unexplained ankle or hindfoot swelling, mass, or atrophy²
 - 8. Patients for whom diagnostic or therapeutic arthroscopy is planned³
 - 9. Patients with recurrent, residual, or new symptoms following ankle surgery³ [34,126,128-131]
 - 10. Patients or relatives with familial hypercholesterolemia or hyperlipidemia [31,132,133]
- 2 Conditions in which intravenous (IV) contrast may be useful
- Conditions in which intra-articular contrast (performed by direct intra-articular injection or indirect joint opacification following IV administration) may be useful

III. QUALIFICATIONS AND RESPONSIBILITIES OF PERSONNEL

See the ACR Practice Parameter for Performing and Interpreting Magnetic Resonance Imaging (MRI) [134].

IV. SAFETY GUIDELINES AND POSSIBLE CONTRAINDICATIONS

See the <u>ACR Practice Parameter for Performing and Interpreting Magnetic Resonance Imaging (MRI)</u> [134] and the <u>ACR Manual on MR Safety</u> [135].

Peer-reviewed literature pertaining to MR safety should be reviewed on a regular basis [136,137].

V. SPECIFICATIONS OF THE EXAMINATION

The written or electronic request for MRI of the ankle and hindfoot should provide sufficient information to demonstrate the medical necessity of the examination and allow for its proper performance and interpretation. Documentation that satisfies medical necessity includes 1) signs and symptoms and/or 2) relevant history (including known diagnoses). Additional information regarding the specific reason for the examination or a provisional diagnosis would be helpful and may at times be needed to allow for the proper performance and interpretation of the examination.

The request for the examination must be originated by a physician or other appropriately licensed health care provider. The accompanying clinical information should be provided by a physician or other appropriately licensed health care provider familiar with the patient's clinical problem or question and consistent with the state's scope of practice requirements. (ACR Resolution 35 adopted in 2006 – revised in 2016, Resolution 12-b) The supervising physician must have adequate understanding of the indications, risks, and benefits of the imaging examination as well as alternative imaging procedures. The physician must be familiar with potential hazards associated with MRI, including potential adverse reactions to contrast media (see the <u>ACR Manual on Contrast Media</u> [138]). The physician should be familiar with relevant ancillary studies that the patient may have

undergone. The physician performing MRI interpretation must have a clear understanding and knowledge of the anatomy and pathophysiology relevant to the MRI examination.

The supervising physician must also understand the pulse sequences to be used and their effect on the appearance of the images, including the potential generation of image artifacts. Standard imaging protocols may be established and varied on a case-by-case basis when necessary. These protocols should be reviewed and updated periodically.

A. Patient Selection

The physician responsible for the examination should supervise patient selection and preparation and be available in person or by telephone for consultation. Patients must be screened and interviewed prior to the examination to exclude those who may be at risk by exposure to the MR environment.

Certain indications require administration of intravenous (IV) contrast media. IV contrast enhancement should be performed using appropriate injection protocols and in accordance with the institution's policy on IV contrast utilization (see the <u>ACR–SPR Practice Parameter for the Use of Intravascular Contrast Media</u> [139]).

Pediatric patients or patients suffering from anxiety or claustrophobia may require sedation or additional assistance. Administration of moderate sedation or general anesthesia may be needed to achieve a successful examination, particularly in young children. If moderate sedation is necessary, refer to the <u>ACR–SIR Practice Parameter for Minimal and/or Moderate Sedation/Analgesia</u> [140].

B. Facility Requirements

Appropriate emergency equipment and medications must be immediately available to treat adverse reactions associated with administered medications. The equipment and medications should be monitored for inventory and drug expiration dates on a regular basis. The equipment, medications, and other emergency support must also be appropriate for the range of ages and sizes in the patient population.

C. Examination Technique

Diagnostic-quality ankle and hindfoot MRI is possible using a variety of magnet designs (closed or open) and field strengths [112]. Utilizing a 3T scanner can provide multiple advantages over 1T or 1.5T. The higher field strength results in a near linear increase in signal-to-noise ratio (SNR), which can be used to improve resolution, decrease acquisition time, or both. However, this comes at the cost of increased metallic susceptibility with resulting greater signal loss, geometric distortion, and heterogeneous fat suppression as well as more pronounced chemical shift and pulsation artifacts [141,142]. Additionally, at least in cadaveric ankle studies, images obtained with a 3T scanner may have higher accuracy for articular cartilage abnormalities in the ankle compared with those obtained with 1T or 1.5T scanners [143,144]. When imaging the ankle and hindfoot on low-field scanners, the reduced SNR necessitates modifications in the imaging parameters [146]. For example, the number of signals averaged can be increased at the expense of longer imaging times and increased risk of involuntary patient motion [147]. Alternatively, the voxel size can be increased (by a combination of larger field of view [FOV], thicker slices, and/or decreased matrix) at the expense of spatial resolution. In addition, fat-suppression techniques that rely on the difference between fat and water precessional frequencies (chemical shift) are unreliable at low field strength and substituting short-tau inversion recovery (STIR) images may be necessary.

MR scanners that are built for extremity-only imaging can also be used for the hindfoot and ankle [120,148]. Low-field—dedicated extremity machines are more susceptible to artifacts and degraded image quality than their high-field counterparts [148,149]. Conversely, newer high-field—dedicated scanners impose a limitation on the useable FOV, which may make imaging a long structure like the entire Achilles tendon problematic.

Regardless of system design, a local receiver coil is mandatory to maximize the SNR [150]. Several choices are available for the ankle and hindfoot [151]. Although not widely available, a local gradient coil can be used to generate images with extremely high resolution to depict the fine detail of anatomic structures like the ankle collateral ligaments [145]. A whole-volume extremity coil allows examination of the ankle in

neutral position or plantar flexion, with the patient lying supine or possibly prone, if the hardware allows. A pair of surface coils joined in an array or in a Helmholtz configuration can substitute for a whole-volume coil, if one is not available. If both ankles are to be imaged, examining each side separately with a smaller coil (eg, an extremity coil) will provide better SNR and higher-resolution imaging [152] than can be achieved by imaging both ankles together in a larger coil (eg, a head coil). A single surface coil can image superficial structures, but for high-resolution imaging of very small, relatively superficial structures, a microscopy coil will provide the SNR necessary at the expense of anatomic coverage [96]. Newer multichannel coils containing multiple coil elements will further increase SNR and are required in use of techniques like parallel imaging that decrease the time of the scan. The coil selected for a given study will also influence limb positioning.

Patient and hindfoot positioning may be individually tailored to the specific indication(s). Allowing the patient to plantar flex the ankle avoids aliasing of the toes onto the heel when the phase direction is oriented along the long axis of the foot [151]. Plantar flexion, which is possible with the patient either supine or prone, also reorients the medial and lateral ankle tendons so that a single imaging plane can show a larger length of each in cross section and so that a smaller segment of each tendon will pass through the magic angle [153,154]. However, this nonstandard position may make visualization of the ankle ligaments more difficult [54,155] and may make it harder to include the entire Achilles tendon in the FOV. The prone position is more comfortable for some patients, reduces involuntary motion, and may reduce claustrophobic feelings in susceptible individuals [156].

Ankle and hindfoot MRI usually includes images acquired in both the short axis and long axis of the foot. These may be sagittal, coronal, transverse, or oblique to the bore of the magnet and to the limb, depending on the position of the ankle. It is beneficial to orient the imaging planes orthogonal to a specific anatomic structure: for the ankle, the talar dome is commonly used; for the hindfoot, the posterior subtalar joint is used [37]. Multiplanar images can be acquired directly or reconstructed electronically from volumetric data acquired in one imaging plane. Standard MR software also allows the prescription of oblique images in virtually any plane if images oriented along the course of a given structure are needed [157]. Oblique sagittal, axial, and coronal sequences have been utilized for improved assessment of the peroneal tendons, syndesmotic ligaments, and calcaneofibular ligaments, respectively, although are not required [141]. The size of the anatomic structures under consideration and the suspected pathology determine the necessary FOV. For example, visualization of the entire extent of a large Achilles tendon tear may require a 22-cm FOV in the sagittal plane, although a FOV of 12 cm or smaller in the coronal plane may be needed to demonstrate small chondral defects around the ankle joint, the syndesmotic ligaments, and the contents of the sinus tarsi [55,143,144]. For routine ankle and hindfoot studies, a FOV of 16 cm or less is desirable for detecting most clinically relevant disorders. A rectangular FOV for coronal and transaxial images of the hindfoot can save imaging time without sacrificing in-plane resolution [151]. Slice thickness should be 4 mm or less to minimize partial-volume effects, but thinner sections may be advantageous for detailed analysis of the ligaments and articular cartilage. Typically, an interslice gap no wider than 10% of the slice width will ensure complete visualization of the intra-articular structures. However, when relatively larger abnormalities like tumors or infections are being imaged, an interslice gap of up to 33% may be useful to increase anatomic coverage and/or decrease imaging time in studies in which motion artifacts are unavoidable. Acquiring 3-D gradient-recalled images as a volume and reconstructing them into 1-mm or thinner slices can assist in the visualization of normal and abnormal ligaments and intratendinous disorders [158], although multiplanar reconstructions of the ankle ligaments can also be produced from conventional 2-D sequences [159]. 3-D imaging is currently used in many centers using turbo spin-echo (TSE)—type sequence (SPACE, VISTA, etc), which allows sub-1-mm scanning. The imaging matrix should balance SNR with desired in-plane spatial resolution and reduction of truncation artifacts, but should be at least 192 steps in the phase-encoding direction and 256 steps in the frequency-encoding direction for 2-D imaging. Even higher matrices combined with smaller FOVs can show fine intratendinous detail [30,160]. A wide variety of pulse sequences—conventional spin-echo, fast (turbo) spin-echo, and gradient-recalled echo—are available for ankle and hindfoot MRI [151]. The choice of sequences may be optimized to address specific clinical questions, and many practices tailor protocols based on the suspected pathology. A typical imaging protocol will be composed of 1 or more pulse sequence types. The exact repetition time (TR), echo time (TE), and flip angle chosen will depend on the field strength of the magnet and the desired relative contrast weighting.

IV contrast enhancement may be useful for evaluating ankle and hindfoot tumors and infections [98,100,101] and may have an adjunct role for tendon imaging [32] and for intrasynovial disorders [60].

Additionally, gadolinium-enhanced MR arthrography, by either direct intra-articular injection or indirect diffusion into the joint following IV injection, may improve diagnostic performance for detecting chronic ligament abnormalities, osteochondral and articular cartilage defects, and impingement lesions [13,29,51,71,95]. Spin-echo, fast spin-echo, or gradient-recalled T1-weighted images with fat suppression are used for contrast-enhanced MR sequences [52,60] At least 1 fluid-sensitive sequence is still necessary when performing MR arthrography to detect extra-articular pathology, as well as at least 1 T1-weighted sequence without fat suppression for evaluating bone marrow and characterizing soft-tissue lesions. Suppressing the signal from fat may enhance the diagnostic yield of some pulse sequences [152]. Fat suppression is most frequently performed using spectrally selective radiofrequency (RF) pulses; however, this technique is limited by field heterogeneity. Dixon techniques are widely available and provide superior fat signal suppression in the setting of field heterogeneity accentuated by the off-center position and irregular air-soft-tissue interfaces often encountered when imaging the ankle and foot [167]. Other methods that are less reliant on field heterogeneity include pulse sequences that decompose fat and water iteratively or null signal from short T1 tissues, such as fat using an inversion pulse (STIR) [168-172]. The STIR technique may be necessary on low-field systems but will also null gadolinium on gadolinium contrast acquisitions. Techniques that rely on separate acquisitions to obtain separate fat and water images are prone to misregistration artifacts because of motion, but combining these sequences with a motioncorrection algorithm can result in robust fat suppression in reasonable scan times [173]. Fat suppression is a useful adjunct to T1-weighted images when IV contrast is used or when MR arthrography is performed with a dilute gadolinium mixture [65,174].

It may be possible to shorten the time required for an ankle or hindfoot MR examination without compromising diagnostic yield. Parallel imaging techniques decrease acquisition times for individual pulse sequences, but at the expense of decreased SNR and the required use of a multichannel receiver coil [175,176]. Alternatively, newer fast 3-D gradient-recalled and fast spin-echo sequences can produce near-isotropic images that can be reconstructed into multiple imaging planes; using these methods, a single volumetric acquisition can substitute for several acquisitions in separate imaging planes, thereby decreasing the total time required for a complete examination [60,174].

Various techniques are useful to minimize artifacts that can degrade image quality. Aliasing is reduced or eliminated by repositioning the extremity (eg, plantar flexing the hindfoot to prevent images of the toes from superimposing on the heel), by orienting the phase-encoding direction anterior-to-posterior when possible, by shielding body parts outside of the area of interest, or by the use of phase oversampling [151,177,178]. Gentle immobilization combined with patient comfort measures best controls involuntary motion [152], although newer pulse sequences can partly correct for some limb motion [173]. Presaturation pulses or gradient moment nulling will reduce ghosting artifacts from flowing blood and other periodic motion [177,179]. Chemical shift artifact is most severe at high field strengths and may necessitate an increase in the receiver bandwidth on high-field scanners [147,177]. Susceptibility artifacts, which originate from heterogeneity of the local field, are also more severe at higher field strengths, in the presence of metallic implants, and when using gradient-recalled pulse sequences. Avoiding gradient-echo imaging and reducing the voxel size by increasing the imaging matrix and/or decreasing the slice thickness and FOV will help reduce the magnitude of susceptibility artifacts [177]. Lastly, magic angle artifact can produce apparently increased signal intensity on short TE images within tendons that curve around the ankle, mimicking intratendinous pathology [180-182]. Plantar flexing the hindfoot to reorient the tendons can reduce this phenomenon [154]. Confirming abnormal signal intensity in the tendons on images with a longer TE and correlating apparent signal intensity abnormalities with changes in tendon thickness will also help avoid this pitfall [151,153].

It is the responsibility of the supervising physician to determine whether additional or unconventional pulse sequences and imaging techniques confer added benefit for the diagnosis and management of the patient. Examinations that use techniques not approved by the Food and Drug Administration, such as the intra-articular injection of gadolinium chelates (direct MR arthrography) [183], can be considered when they are judged to be medically appropriate.

VI. DOCUMENTATION

Reporting should be in accordance with the <u>ACR Practice Parameter for Communication of Diagnostic Imaging Findings</u> [184].

At a minimum, the report should address the condition of the major ankle tendons, ligaments, and joints. In selected cases, a description of findings in the bone and bone marrow, synovium, joints, retinacula, muscles,

sinus tarsi, plantar fascia, neurovascular structures, and subcutaneous tissue would be appropriate. The report should use standard anatomic nomenclature and precise terms for describing identified abnormalities whenever possible.

Specific policies and procedures related to MRI safety should be in place along with documentation that is updated annually and compiled under the supervision and direction of the supervising MRI physician. Guidelines should be provided that deal with potential hazards associated with the MRI examination of the patient as well as to others in the immediate area [136,137,185]. Screening forms must also be provided to detect those patients who may be at risk for adverse events associated with the MRI examination [186].

VII. EQUIPMENT SPECIFICATIONS

Equipment monitoring should be in accordance with the <u>ACR-AAPM Technical Standard for Diagnostic Medical Physics Performance Monitoring of Magnetic Resonance (MR) Imaging Equipment</u> [187].

The MRI equipment specifications and performance must meet all state and federal requirements. The requirements include, but are not limited to, specifications of maximum static magnetic strength, maximum rate of change of the magnetic field strength (dB/dt), maximum radiofrequency power deposition (specific absorption rate), and maximum acoustic noise levels.

VIII. QUALITY CONTROL AND IMPROVEMENT, SAFETY, INFECTION CONTROL, AND PATIENT EDUCATION

Policies and procedures related to quality, patient education, infection control, and safety should be developed and implemented in accordance with the ACR Policy on Quality Control and Improvement, Safety, Infection Control, and Patient Education appearing under the heading ACR Position Statement on Quality Control & Improvement, Safety, Infection Control, and Patient Education on the ACR website (https://www.acr.org/Advocacy-and-Economics/ACR-Position-Statements/Quality-Control-and-Improvement).

ACKNOWLEDGEMENTS

This practice parameter was developed according to the process described under the heading *The Process for Developing ACR Practice Parameters and Technical Standards* on the ACR website (https://www.acr.org/Clinical-Resources/Practice-Parameters-and-Technical-Standards) by the Committee on Practice Parameters – Body Imaging (Musculoskeletal) of the ACR Commission on Body Imaging and the Committee on Practice Parameters – Pediatric Radiology of the ACR Commission on Pediatric Radiology, in collaboration with the SPR, and the SSR. Writing Committee – members represent their societies in the initial and final revision of this practice parameter

ACR SPR SSR

Bethany U. Casagranda, DO, Chair Delma Jarrett, MD Tetyana Gorbachova, MD

Jeffrey M. Brody, MD, FACR Amisha Shah, MD

Joseph A. Delic, MD Jason Higgins, MD

Committee on Body Imaging (Musculoskeletal)

(ACR Committee responsible for sponsoring the draft through the process)

Catherine C. Roberts, MD, Chair Suzanne S. Long, MD

Jeffrey M. Brody, MD, FACR Kambiz Motamedi, MD

Bethany U. Casagranda, DO Carlos A. Rivera, BSc

Elaine S Gould, MD, FACR Aleksandr Rozenberg, MD

Mary K. Jesse, MD Naveen Subhas, MD

Kenneth S. Lee, MD

Committee on Practice Parameters - Pediatric Radiology

(ACR Committee responsible for sponsoring the draft through the process)

Terry L. Levin, MD, FACR, Chair

Jane Sun Kim, MD

John B. Amodio, MD, FACR

Jennifer A Knight, MD

Jessica Kurian, MD

Committee on Practice Parameters - Pediatric Radiology

Tara M. Catanzano, MB, BCh Matthew P. Lungren, MD, MPH

Harris L. Cohen, MD, FACR Helen R. Nadel, MD Kassa Darge, MD, PhD Erica Poletto, MD

Dorothy L. Gilbertson-Dahdal, MD Richard B. Towbin, MD, FACR

Lauren P. Golding, MD

Safwan S. Halabi, MD

Esben S. Vogelius, MD

Jason Higgins, DO

Andrew B. Rosenkrantz, MD, Chair, Commission on Body Imaging Richard A. Barth, MD, FACR, Chair, Commission on Pediatric Radiology David B. Larson, MD, MBA, Chair, Commission on Quality and Safety

Mary S. Newell, MD, FACR, Chair, Committee on Practice Parameters and Technical Standards

Comment Reconciliation Committee

Eric Friedberg, MD, FACR, Chair Jason Higgins, MD
Rachel Gerson, MD, Co-Chair Delma Jarrett, MD

Richard A. Barth, MD, FACR

Nicholas M. Beckmann, MD

David B. Larson, MD, MBA

Jeffrey M. Brody, MD, FACR

Bethany U. Casagranda, DO

Avneesh Chhabra, MD

Joseph A. Delic, MD

Richard Duszak Jr., MD, FACR

Amy Kotsenas, MD, FACR

David B. Larson, MD, MBA

Paul A. Larson, MD, FACR

Terry L. Levin, MD, FACR

Mary S. Newell, MD, FACR

Catherine C. Roberts, MD

Andrew B. Rosenkrantz, MD

Tetyana Gorbachova, MD Amisha Shah, MD

REFERENCES

- **1.** Anzilotti KK, Schweitzer MM, Hecht PP, Wapner KK, Kahn MM, Ross MM. Effect of foot and ankle MR imaging on clinical decision making. Radiology 201:515-7, .
- **2.** Bearcroft PP, Guy SS, Bradley MM, Robinson FF. MRI of the ankle: effect on diagnostic confidence and patient management. AJR Am J Roentgenol 187:1327-31, .
- **3.** Conti SS, Michelson JJ, Jahss MM. Clinical significance of magnetic resonance imaging in preoperative planning for reconstruction of posterior tibial tendon ruptures. Foot Ankle 13:208-14, .
- **4.** Greenstein AS, Marzo-Ortega H, Emery P, O'Connor P, McGonagle D. Magnetic resonance imaging as a predictor of progressive joint destruction in neuropathic joint disease. Arthritis Rheum. 2002 Oct;46(10):2814-5.
- **5.** Nicholson CC, Berlet GG, Lee TT. Prediction of the success of nonoperative treatment of insertional Achilles tendinosis based on MRI. Foot Ankle Int 28:472-7, .
- **6.** Takao MM, Innami KK, Matsushita TT, Uchio YY, Ochi MM. Arthroscopic and magnetic resonance image appearance and reconstruction of the anterior talofibular ligament in cases of apparent functional ankle instability. Am J Sports Med 36:1542-7, .
- **7.** Zanetti MM, De Simoni CC, Wetz HH, Zollinger HH, Hodler JJ. Magnetic resonance imaging of injuries to the ankle joint: can it predict clinical outcome?. Skeletal Radiol 26:82-8, .
- **8.** Tocci SL, Madom IA, Bradley MP, Langer PR, DiGiovanni CW. The diagnostic value of MRI in foot and ankle surgery. Foot Ankle Int. 28(2):166-8, 2007 Feb.
- **9.** Groshar DD, Gorenberg MM, Ben-Haim SS, Jerusalmi JJ, Liberson AA. Lower extremity scintigraphy: the foot and ankle. Semin Nucl Med 28:62-77, .
- **10.** Cooper AA, Mizel MM, Patel PP, Steinmetz NN, Clifford PP. Comparison of MRI and local anesthetic tendon sheath injection in the diagnosis of posterior tibial tendon tenosynovitis. Foot Ankle Int 28:1124-7, .
- **11.** Newman JS.. Diagnostic and therapeutic injections of the foot and ankle. [Review] [21 refs]. Semin Roentgenol. 39(1):85-94, 2004 Jan.
- 12. Sugimoto K, Takakura Y, Samoto N, Nakayama S, Tanaka Y. Subtalar arthrography in recurrent instability of the

- ankle. Clin Orthop Relat Res. 2002 Jan; (394):169-76.
- **13.** Taljanovic MS, Alcala JN, Gimber LH, Rieke JD, Chilvers MM, Latt LD. High-resolution US and MR imaging of peroneal tendon injuries. [Review][Erratum appears in Radiographics. 2015 Mar-Apr;35(2):651; PMID: 25763746]. Radiographics. 35(1):179-99, 2015 Jan-Feb.
- **14.** Sconfienza LM, Orlandi D, Lacelli F, Serafini G, Silvestri E. Dynamic high-resolution US of ankle and midfoot ligaments: normal anatomic structure and imaging technique. [Review]. Radiographics. 35(1):164-78, 2015 Jan-Feb.
- **15.** Khoury V, Guillin R, Dhanju J, Cardinal E. Ultrasound of ankle and foot: overuse and sports injuries. [Review] [119 refs]. Semin Musculoskelet Radiol. 11(2):149-61, 2007 Jun.
- 16. Schreibman KK. Ankle tenography: what, how, and why. Semin Roentgenol 39:95-113, .
- **17.** Oae K, Takao M, Uchio Y, Ochi M. Evaluation of anterior talofibular ligament injury with stress radiography, ultrasonography and MR imaging. Skeletal Radiol. 2010;39(1):41-47.
- **18.** Cutler L, Molloy A, Dhukuram V, Bass A. Do CT scans aid assessment of distal tibial physeal fractures?. J Bone Joint Surg Br. 86(2):239-43, 2004 Mar.
- **19.** Haapamaki VV, Kiuru MJ, Koskinen SK. Ankle and foot injuries: analysis of MDCT findings. AJR Am J Roentgenol. 183(3):615-22, 2004 Sep.
- **20.** Hochman M, Reed MH. Features of calcaneonavicular coalition on coronal computed tomography. Skeletal Radiol. 2000 Jul;29(7):409-12.
- **21.** Linsenmaier U, Brunner U, Schöning A, et al. Classification of calcaneal fractures by spiral computed tomography: implications for surgical treatment. Eur Radiol. 2003 Oct;13(10):2315-22.
- **22.** Rubino RR, Valderrabano VV, Sutter PP, Regazzoni PP. Prognostic value of four classifications of calcaneal fractures. Foot Ankle Int 30:229-38, .
- **23.** Verhagen RA, Maas M, Dijkgraaf MG, Tol JL, Krips R, van Dijk CN. Prospective study on diagnostic strategies in osteochondral lesions of the talus. Is MRI superior to helical CT? J Bone Joint Surg Br. 2005;87(1):41-46.
- **24.** Kiss ZZ, Khan KK, Fuller PP. Stress fractures of the tarsal navicular bone: CT findings in 55 cases. AJR Am J Roentgenol 160:111-5, .
- **25.** Rosenberg ZZ, Jahss MM, Noto AA, et al. Rupture of the posterior tibial tendon: CT and surgical findings. Radiology 167:489-93, .
- **26.** Choplin RH, Buckwalter KA, Rydberg J, Farber JM. CT with 3D rendering of the tendons of the foot and ankle: technique, normal anatomy, and disease. [Review] [13 refs]. Radiographics. 24(2):343-56, 2004 Mar-Apr.
- **27.** Pelc JS, Beaulieu CF. Volume rendering of tendon-bone relationships using unenhanced CT. AJR Am J Roentgenol. 2001 Apr;176(4):973-7.
- **28.** Schmid MR, Pfirrmann CW, Hodler J, Vienne P, Zanetti M. Cartilage lesions in the ankle joint: comparison of MR arthrography and CT arthrography. Skeletal Radiol. 2003 May;32(5):259-65.
- **29.** Karjalainen PT, Soila K, Aronen HJ, et al. MR imaging of overuse injuries of the Achilles tendon. AJR Am J Roentgenol. 2000 Jul;175(1):251-60.
- **30.** Koivunen-Niemelä TT, Komu MM, Viikari JJ, Alanen AA. Magnetic resonance imaging of Achilles tendon xanthomas using a fat-water discrimination technique at 0.1 T. Acad Radiol 2:319-23, .
- **31.** Marshall H, Howarth C, Larkman DJ, Herlihy AH, Oatridge A, Bydder GM. Contrast-enhanced magic-angle MR imaging of the Achilles tendon. AJR Am J Roentgenol. 2002 Jul;179(1):187-92.
- **32.** Schweitzer ME, Karasick D. MR imaging of disorders of the Achilles tendon. AJR Am J Roentgenol. 2000 Sep;175(3):613-25.
- **33.** Shalabi A, Kristoffersen-Wiberg M, Aspelin P, Movin T. MR evaluation of chronic Achilles tendinosis. A longitudinal study of 15 patients preoperatively and two years postoperatively. Acta Radiol. 2001 May;42(3):269-76.
- **34.** Bencardino JJ, Rosenberg ZZ, Beltran JJ, et al. MR imaging of dislocation of the posterior tibial tendon. AJR Am J Roentgenol 169:1109-12, .
- **35.** Bouysset M, Tebib J, Tavernier T, et al. Posterior tibial tendon and subtalar joint complex in rheumatoid arthritis: magnetic resonance imaging study. J Rheumatol. 2003 Sep;30(9):1951-4.
- **36.** Khoury NN, el-Khoury GG, Saltzman CC, Brandser EE. MR imaging of posterior tibial tendon dysfunction. AJR Am J Roentgenol 167:675-82, .
- **37.** Nallamshetty L, Nazarian LN, Schweitzer ME, et al. Evaluation of posterior tibial pathology: comparison of sonography and MR imaging. Skeletal Radiol. 2005;34(7):375-380.
- **38.** Premkumar A, Perry MB, Dwyer AJ, et al. Sonography and MR imaging of posterior tibial tendinopathy. AJR Am J Roentgenol. 2002 Jan;178(1):223-32.
- 39. Donovan AA, Rosenberg ZZ. Extraarticular lateral hindfoot impingement with posterior tibial tendon tear: MRI

correlation. AJR Am J Roentgenol 193:672-8, .

- **40.** Khoury NN, el-Khoury GG, Saltzman CC, Kathol MM. Peroneus longus and brevis tendon tears: MR imaging evaluation. Radiology 200:833-41, .
- **41.** Rademaker J, Rosenberg ZS, Delfaut EM, Cheung YY, Schweitzer ME. Tear of the peroneus longus tendon: MR imaging features in nine patients. Radiology. 2000 Mar;214(3):700-4.
- **42.** Rosenberg ZZ, Beltran JJ, Cheung YY, Colon EE, Herraiz FF. MR features of longitudinal tears of the peroneus brevis tendon. AJR Am J Roentgenol 168:141-7, .
- **43.** Rosenberg ZZ, Bencardino JJ, Astion DD, Schweitzer MM, Rokito AA, Sheskier SS. MRI features of chronic injuries of the superior peroneal retinaculum. AJR Am J Roentgenol 181:1551-7, .
- **44.** Beischer AA, Beamond BB, Jowett AA, O'Sullivan RR. Distal tendinosis of the tibialis anterior tendon. Foot Ankle Int 30:1053-9, .
- **45.** Bouysset MM, Tavernier TT, Tebib JJ, et al. CT and MRI evaluation of tenosynovitis of the rheumatoid hindfoot. Clin Rheumatol 14:303-7, .
- **46.** Gallo RR, Kolman BB, Daffner RR, Sciulli RR, Roberts CC, DeMeo PP. MRI of tibialis anterior tendon rupture. Skeletal Radiol 33:102-6, .
- **47.** Lo LD, Schweitzer ME, Fan JK, Wapner KL, Hecht PJ. MR imaging findings of entrapment of the flexor hallucis longus tendon. AJR Am J Roentgenol. 2001 May;176(5):1145-8.
- **48.** Mengiardi BB, Pfirrmann CC, Vienne PP, et al. Anterior tibial tendon abnormalities: MR imaging findings. Radiology 235:977-84, .
- **49.** Cerezal L, Llopis E, Canga A, Rolón A. MR arthrography of the ankle: indications and technique. Radiol Clin North Am. 2008 Nov;46(6):973-94, v.
- **50.** Brown KK, Morrison WW, Schweitzer MM, Parellada JJ, Nothnagel HH. MRI findings associated with distal tibiofibular syndesmosis injury. AJR Am J Roentgenol 182:131-6, .
- **51.** Chandnani VV, Harper MM, Ficke JJ, et al. Chronic ankle instability: evaluation with MR arthrography, MR imaging, and stress radiography. Radiology 192:189-94, .
- **52.** Kim SS, Huh YY, Song HH, et al. Chronic tibiofibular syndesmosis injury of ankle: evaluation with contrast-enhanced fat-suppressed 3D fast spoiled gradient-recalled acquisition in the steady state MR imaging. Radiology 242:225-35, .
- **53.** Klein MM. MR imaging of the ankle: normal and abnormal findings in the medial collateral ligament. AJR Am J Roentgenol 162:377-83, .
- **54.** Mesgarzadeh MM, Schneck CC, Tehranzadeh JJ, Chandnani VV, Bonakdarpour AA. Magnetic resonance imaging of ankle ligaments. Emphasis on anatomy and injuries to lateral collateral ligaments. Magn Reson Imaging Clin N Am 2:39-58, .
- **55.** Oae K, Takao M, Naito K, et al. Injury of the tibiofibular syndesmosis: value of MR imaging for diagnosis. Radiology. 2003 Apr;227(1):155-61.
- **56.** Toye LR, Helms CA, Hoffman BD, Easley M, Nunley JA. MRI of spring ligament tears. AJR Am J Roentgenol. 184(5):1475-80, 2005 May.
- **57.** Uys HD, Rijke AM. Clinical association of acute lateral ankle sprain with syndesmotic involvement: a stress radiography and magnetic resonance imaging study. Am J Sports Med. 2002;30(6):816-22.
- **58.** Crim J, Longenecker LG. MRI and surgical findings in deltoid ligament tears. AJR Am J Roentgenol. 204(1):W63-9, 2015 Jan.
- **59.** Bureau NJ, Cardinal E, Hobden R, Aubin B. Posterior ankle impingement syndrome: MR imaging findings in seven patients. Radiology. 2000 May;215(2):497-503.
- **60.** Choo HH, Suh JJ, Kim SS, Huh YY, Kim MM, Lee JJ. Ankle MRI for anterolateral soft tissue impingement: increased accuracy with the use of contrast-enhanced fat-suppressed 3D-FSPGR MRI. Korean J Radiol 9:409-15, .
- **61.** Duncan D, Mologne T, Hildebrand H, Stanley M, Schreckengaust R, Sitler D. The usefulness of magnetic resonance imaging in the diagnosis of anterolateral impingement of the ankle. J Foot Ankle Surg. 2006;45(5):304-307.
- **62.** Lee JJ, Calder JJ, Healy JJ. Posterior impingement syndromes of the ankle. Semin Musculoskelet Radiol 12:154-69.
- **63.** Messiou C, Robinson P, O'Connor PJ, Grainger A. Subacute posteromedial impingement of the ankle in athletes: MR imaging evaluation and ultrasound guided therapy. Skeletal Radiol. 35(2):88-94, 2006 Feb.
- **64.** Robinson P, White LM, Salonen D, Ogilvie-Harris D. Anteromedial impingement of the ankle: using MR arthrography to assess the anteromedial recess. AJR Am J Roentgenol. 2002 Mar;178(3):601-4.
- **65.** Robinson P, White LM, Salonen DC, Daniels TR, Ogilvie-Harris D. Anterolateral ankle impingement: mr arthrographic assessment of the anterolateral recess. Radiology. 2001 Oct;221(1):186-90.

- **66.** Subhas NN, Vinson EE, Cothran RR, Santangelo JJ, Nunley JJ, Helms CC. MRI appearance of surgically proven abnormal accessory anterior-inferior tibiofibular ligament (Bassett's ligament). Skeletal Radiol 37:27-33, .
- **67.** Donovan A, Rosenberg ZS. MRI of ankle and lateral hindfoot impingement syndromes. [Review] [74 refs]. AJR Am J Roentgenol. 195(3):595-604, 2010 Sep.
- **68.** Bui-Mansfield LT, Kline M, Chew FS, Rogers LF, Lenchik L. Osteochondritis dissecans of the tibial plafond: imaging characteristics and a review of the literature. AJR Am J Roentgenol. 2000 Nov;175(5):1305-8.
- **69.** Lee KB, Bai LB, Park JG, Yoon TR. A comparison of arthroscopic and MRI findings in staging of osteochondral lesions of the talus. Knee Surg Sports Traumatol Arthrosc. 2008;16(11):1047-1051.
- **70.** Magee TT, Hinson GG. Usefulness of MR imaging in the detection of talar dome injuries. AJR Am J Roentgenol 170:1227-30, .
- **71.** Naran KN, Zoga AC. Osteochondral lesions about the ankle. [Review] [16 refs]. Radiol Clin North Am. 46(6):995-1002, v, 2008 Nov.
- **72.** Bae SS, Lee HH, Lee KK, et al. Comparison of arthroscopic and magnetic resonance imaging findings in osteochondral lesions of the talus. Foot Ankle Int 33:1058-62, .
- **73.** Griffith JF, Lau DT, Yeung DK, Wong MW. High-resolution MR imaging of talar osteochondral lesions with new classification. Skeletal Radiol. 2012; 41(4):387-399.
- **74.** Gatlin CC, Matheny LL, Ho CC, Johnson NN, Clanton TT. Diagnostic accuracy of 3.0 Tesla magnetic resonance imaging for the detection of articular cartilage lesions of the talus. Foot Ankle Int 36:288-92, .
- **75.** Allen JJ, Greer BB, Sorge DD, Campbell SS. MR imaging of neuropathies of the leg, ankle, and foot. Magn Reson Imaging Clin N Am 16:117-31, vii, .
- **76.** Chung KW, Suh BC, Shy ME, et al. Different clinical and magnetic resonance imaging features between Charcot-Marie-Tooth disease type 1A and 2A. Neuromuscul Disord. 18(8):610-8, 2008 Aug.
- **77.** Delfaut EM, Demondion X, Bieganski A, Thiron MC, Mestdagh H, Cotten A. Imaging of foot and ankle nerve entrapment syndromes: from well-demonstrated to unfamiliar sites. Radiographics. 2003;23(3):613-23.
- 78. Finkel JJ. Tarsal tunnel syndrome. Magn Reson Imaging Clin N Am 2:67-78, .
- **79.** Frey C, Kerr R. Magnetic resonance imaging and the evaluation of tarsal tunnel syndrome. Foot Ankle. 1993;14(3):159-164.
- **80.** Dong Q, Jacobson JA, Jamadar DA, et al. Entrapment neuropathies in the upper and lower limbs: anatomy and MRI features. Radiol Res Pract. 2012;2012:230679.
- **81.** Grasel RP, Schweitzer ME, Kovalovich AM, et al. MR imaging of plantar fasciitis: edema, tears, and occult marrow abnormalities correlated with outcome. AJR Am J Roentgenol. 1999 Sep;173(3):699-701.
- **82.** Jeswani T, Morlese J, McNally EG. Getting to the heel of the problem: plantar fascia lesions. [Review] [26 refs]. Clin Radiol. 64(9):931-9, 2009 Sep.
- **83.** Yu JS. Pathologic and post-operative conditions of the plantar fascia: review of MR imaging appearances. Skeletal Radiol. 2000 Sep;29(9):491-501.
- **84.** Yu JS, Spigos D, Tomczak R. Foot pain after a plantar fasciotomy: an MR analysis to determine potential causes. J Comput Assist Tomogr. 1999;23(5):707-12.
- **85.** Lektrakul N, Chung CB, Lai Ym, et al. Tarsal sinus: arthrographic, MR imaging, MR arthrographic, and pathologic findings in cadavers and retrospective study data in patients with sinus tarsi syndrome. Radiology. 2001 Jun;219(3):802-10.
- **86.** Bottger BB, Schweitzer MM, El-Noueam KK, Desai MM. MR imaging of the normal and abnormal retrocalcaneal bursae. AJR Am J Roentgenol 170:1239-41, .
- **87.** Maillefert JF, Dardel P, Cherasse A, Mistrih R, Krause D, Tavernier C. Magnetic resonance imaging in the assessment of synovial inflammation of the hindfoot in patients with rheumatoid arthritis and other polyarthritis. Eur J Radiol. 2003 Jul;47(1):1-5.
- **88.** Rochwerger A, Groulier P, Curvale G, Launay F. Pigmented villonodular synovitis of the foot and ankle: a report of eight cases. Foot Ankle Int. 1999 Sep;20(9):587-90.
- **89.** Weishaupt D, Schweitzer ME, Morrison WB, Haims AH, Wapner K, Kahn M. MRI of the foot and ankle: prevalence and distribution of occult and palpable ganglia. J Magn Reson Imaging. 2001 Oct;14(4):464-71.
- **90.** Elias I, Zoga AC, Raikin SM, et al. Bone stress injury of the ankle in professional ballet dancers seen on MRI. BMC Musculoskeletal Disorders. 9:39, 2008 Mar 28.
- **91.** Fernandez-Canton G, Casado O, Capelastegui A, Astigarraga E, Larena JA, Merino A. Bone marrow edema syndrome of the foot: one year follow-up with MR imaging. Skeletal Radiol. 2003 May;32(5):273-8.
- **92.** Schmid MR, Hodler J, Vienne P, Binkert CA, Zanetti M. Bone marrow abnormalities of foot and ankle: STIR versus T1-weighted contrast-enhanced fat-suppressed spin-echo MR imaging. Radiology. 2002 Aug;224(2):463-9.
- 93. Sijbrandij ES, van Gils AP, Louwerens JW, de Lange EE. Posttraumatic subchondral bone contusions and

- fractures of the talotibial joint: occurrence of "kissing" lesions. AJR Am J Roentgenol. 2000 Dec;175(6):1707-10.
- **94.** Zanetti M, Steiner CL, Seifert B, Hodler J. Clinical outcome of edema-like bone marrow abnormalities of the foot. Radiology. 2002 Jan;222(1):184-8.
- **95.** Beltran J, Shankman S. MR imaging of bone lesions of the ankle and foot. Magn Reson Imaging Clin N Am. 2001 Aug;9(3):553-66, xi.
- **96.** Lee IS, Choi JA, Oh JH, et al. Microscopy coil for preoperative MRI of small soft-tissue masses of the hand and foot: comparison with conventional surface coil. AJR Am J Roentgenol. 191(6):W256-63, 2008 Dec.
- **97.** Llauger JJ, Palmer JJ, Monill JJ, Franquet TT, Bagué SS, Rosón NN. MR imaging of benign soft-tissue masses of the foot and ankle. Radiographics 18:1481-98, .
- **98.** Maldjian C, Rosenberg ZS. MR imaging features of tumors of the ankle and foot. Magn Reson Imaging Clin N Am. 2001 Aug;9(3):639-57, xii.
- **99.** Ganguly A, Aniq H, Skiadas B. Lumps and bumps around the foot and ankle: an assessment of frequency with ultrasound and MRI. Skeletal Radiol. 42(8):1051-60, 2013 Aug.
- **100.** Ledermann HP, Morrison WB, Schweitzer ME. Pedal abscesses in patients suspected of having pedal osteomyelitis: analysis with MR imaging. Radiology. 2002 Sep;224(3):649-55.
- **101.** Ledermann HP, Morrison WB, Schweitzer ME, Raikin SM. Tendon involvement in pedal infection: MR analysis of frequency, distribution, and spread of infection. AJR Am J Roentgenol. 2002 Oct;179(4):939-47.
- **102.** Godoy-Santos AA, Bordalo-Rodrigues MM, Rosemberg LL, et al. Kager's fat pad inflammation associated with HIV infection and AIDS: MRI findings. Skeletal Radiol 43:1257-62, .
- **103.** Chan VV, Morrison WW, Kavanagh EE. Postoperative infection in the foot and ankle. Semin Musculoskelet Radiol 16:241-53, .
- **104.** Cahuzac JP, Baunin C, Luu S, Estivalezes E, Sales de Gauzy J, Hobatho MC. Assessment of hindfoot deformity by three-dimensional MRI in infant club foot. J Bone Joint Surg Br. 1999 Jan;81(1):97-101.
- **105.** Cheung Y, Rosenberg ZS. MR imaging of the accessory muscles around the ankle. Magn Reson Imaging Clin N Am. 2001 Aug;9(3):465-73, x.
- **106.** Miller TT, Staron RB, Feldman F, Parisien M, Glucksman WJ, Gandolfo LH. The symptomatic accessory tarsal navicular bone: assessment with MR imaging. Radiology. 1995;195(3):849-853.
- **107.** Newman JS, Newberg AH. Congenital tarsal coalition: multimodality evaluation with emphasis on CT and MR imaging. Radiographics. 2000;20(2):321-32; quiz 526-7, 532.
- **108.** Patel CV.. The foot and ankle: MR imaging of uniquely pediatric disorders. [Review] [51 refs]. Magn Reson Imaging Clin N Am. 17(3):539-47, vii, 2009 Aug.
- **109.** Zammit JJ, Singh DD. The peroneus quartus muscle. Anatomy and clinical relevance. J Bone Joint Surg Br 85:1134-7, .
- **110.** Lehtinen AA, Paimela LL, Kreula JJ, Leirisalo-Repo MM, Taavitsainen MM. Painful ankle region in rheumatoid arthritis. Analysis of soft-tissue changes with ultrasonography and MR imaging. Acta Radiol 37:572-7, .
- **111.** Erdem CZ, Tekin NS, Sarikaya S, Erdem LO, Gulec S. MR imaging features of foot involvement in patients with psoriasis. Eur J Radiol. 67(3):521-5, 2008 Sep.
- **112.** Eshed I, Althoff CE, Feist E, et al. Magnetic resonance imaging of hindfoot involvement in patients with spondyloarthritides: comparison of low-field and high-field strength units. Eur J Radiol. 65(1):140-7, 2008 Jan.
- **113.** Weishaupt D, Schweitzer ME, Alam F, Karasick D, Wapner K. MR imaging of inflammatory joint diseases of the foot and ankle. Skeletal Radiol. 1999 Dec;28(12):663-9.
- **114.** Yu JJ, Chung CC, Recht MM, Dailiana TT, Jurdi RR. MR imaging of tophaceous gout. AJR Am J Roentgenol 168:523-7, .
- **115.** Bancroft LW, Peterson JJ, Kransdorf MJ. Imaging of soft tissue lesions of the foot and ankle. [Review] [39 refs]. Radiol Clin North Am. 46(6):1093-103, vii, 2008 Nov.
- **116.** American College of Radiology. ACR-SPR-SSR Practice Parameter for the Performance and Interpretation of Magnetic Resonance Imaging (MRI) of Bone and Soft-Tissue Tumors. Available at https://gravitas.acr.org/PPTS/GetDocumentView?docId=88+&releaseId=2
- **117.** Petit PP, Panuel MM, Faure FF, et al. Acute fracture of the distal tibial physis: role of gradient-echo MR imaging versus plain film examination. AJR Am J Roentgenol 166:1203-6, .
- **118.** Robbins MM, Wilson MM, Sella EE. MR imaging of anterosuperior calcaneal process fractures. AJR Am J Roentgenol 172:475-9, .
- 119. Umans HH, Pavlov HH. Insufficiency fracture of the talus: diagnosis with MR imaging. Radiology 197:439-42, .
- **120.** Nikken JJ, Oei EH, Ginai AZ, et al. Acute ankle trauma: value of a short dedicated extremity MR imaging examination in prediction of need for treatment. Radiology. 2005;234(1):134-142.
- 121. Tochigi YY, Yoshinaga KK, Wada YY, Moriya HH. Acute inversion injury of the ankle: magnetic resonance

imaging and clinical outcomes. Foot Ankle Int 19:730-4, .

- **122.** Anderson SS, Weber MM, Steinbach LL, Ballmer FF. Shoe rim and shoe buckle pseudotumor of the ankle in elite and professional figure skaters and snowboarders: MR imaging findings. Skeletal Radiol 33:325-9, .
- **123.** Dunfee WR, Dalinka MK, Kneeland JB. Imaging of athletic injuries to the ankle and foot. Radiol Clin North Am. 2002 Mar;40(2):289-312, vii.
- **124.** Hillier JC, Peace K, Hulme A, Healy JC. Pictorial review: MRI features of foot and ankle injuries in ballet dancers. [Review] [10 refs]. Br J Radiol. 77(918):532-7, 2004 Jun.
- **125.** Zoga AC, Schweitzer ME. Imaging sports injuries of the foot and ankle. Magn Reson Imaging Clin N Am. 2003 May;11(2):295-310.
- **126.** Alparslan LL, Chiodo CC. Lateral ankle instability: MR imaging of associated injuries and surgical treatment procedures. Semin Musculoskelet Radiol 12:346-58, .
- **127.** Kanamoto TT, Shiozaki YY, Tanaka YY, Yonetani YY, Horibe SS. The use of MRI in pre-operative evaluation of anterior talofibular ligament in chronic ankle instability. Bone Joint Res 3:241-5, .
- **128.** Chien AJ, Jacobson JA, Jamadar DA, Brigido MK, Femino JE, Hayes CW. Imaging appearances of lateral ankle ligament reconstruction. Radiographics. 2004;24(4):999-1008.
- **129.** Liem MM, Zegel HH, Balduini FF, Turner MM, Becker JJ, Caballero-Saez AA. Repair of Achilles tendon ruptures with a polylactic acid implant: assessment with MR imaging. AJR Am J Roentgenol 156:769-73, .
- **130.** Zeiss JJ, Fenton PP, Ebraheim NN, Coombs RR. Magnetic resonance imaging for ineffectual tarsal tunnel surgical treatment. Clin Orthop Relat Res 264:264-6, .
- **131.** Zanetti M, Saupe N, Espinosa N. Postoperative MR imaging of the foot and ankle: tendon repair, ligament repair, and Morton's neuroma resection. [Review] [26 refs]. Semin Musculoskelet Radiol. 14(3):357-64, 2010 Sep.
- **132.** Bude RR, Adler RR, Bassett DD. Diagnosis of Achilles tendon xanthoma in patients with heterozygous familial hypercholesterolemia: MR vs sonography. AJR Am J Roentgenol 162:913-7, .
- **133.** Dussault RR, Kaplan PP, Roederer GG. MR imaging of Achilles tendon in patients with familial hyperlipidemia: comparison with plain films, physical examination, and patients with traumatic tendon lesions. AJR Am J Roentgenol 164:403-7, .
- **134.** American College of Radiology. ACR Practice Parameter for Performing and Interpreting Magnetic Resonance Imaging (MRI). Available at: https://gravitas.acr.org/PPTS/GetDocumentView?docId=146+&releaseId=2
- **135.** American College of Radiology. ACR Committee on MR Safety. 2024 ACR Manual on MR Safety. Available at: https://edge.sitecorecloud.io/americancoldf5f-acrorgf92a-productioncb02-
- 3650/media/ACR/Files/Clinical/Radiology-Safety/Manual-on-MR-Safety.pdf.
- **136.** Shellock FG. Reference Manual for Magnetic Resonance Safety, Implants, and Devices. Playa Del Rey, CA Biomedical Research Publishing Group; 2013.
- **137.** Shellock FG, Crues JV. MR procedures: biologic effects, safety, and patient care. Radiology. 2004 Sep;232(3):635-52.
- **138.** American College of Radiology. ACR Committee on Drugs and Contrast Media. Manual on Contrast Media. Available at: https://www.acr.org/Clinical-Resources/Clinical-Tools-and-Reference/Contrast-Manual.
- **139.** American College of Radiology. ACR–SPR Practice Parameter for the use of Intravascular Contrast Media. Available at https://gravitas.acr.org/PPTS/GetDocumentView?docId=142+&releaseId=2
- **140.** American College of Radiology. ACR—SIR Practice Parameter For Minimal and/or Moderate Sedation/Analgesia. Available at https://gravitas.acr.org/PPTS/GetDocumentView?docId=95+&releaseId=2
- **141.** Siriwanarangsun P, Bae WC, Statum S, Chung CB. Advanced MRI Techniques for the Ankle. [Review]. AJR. American Journal of Roentgenology. 209(3):511-524, 2017 Sep. AJR Am J Roentgenol. 209(3):511-524, 2017 Sep.
- **142.** Ramnath RR. 3T MR imaging of the musculoskeletal system (Part I): considerations, coils, and challenges. Magn Reson Imaging Clin N Am. 2006 Feb;14(1):27-40.
- **143.** Barr CC, Bauer JJ, Malfair DD, et al. MR imaging of the ankle at 3 Tesla and 1.5 Tesla: protocol optimization and application to cartilage, ligament and tendon pathology in cadaver specimens. Eur Radiol 17:1518-28, .
- **144.** Schibany NN, Ba-Ssalamah AA, Marlovits SS, et al. Impact of high field (3.0 T) magnetic resonance imaging on diagnosis of osteochondral defects in the ankle joint. Eur J Radiol 55:283-8, .
- **145.** Crooks LL, Arakawa MM, Hoenninger JJ, McCarten BB, Watts JJ, Kaufman LL. Magnetic resonance imaging: effects of magnetic field strength. Radiology 151:127-33, .
- 146. Erickson SJ. High-resolution imaging of the musculoskeletal system. Radiology. 1997 Dec;205(3):593-618.
- **147.** Hottya GA, Péterfy CG, Uffmann M, et al. Dedicated extremity MR imaging of the foot and ankle. Eur Radiol. 2000;10(3):467-75.
- 148. Verhoek GG, Zanetti MM, Duewell SS, Zollinger HH, Hodler JJ. MRI of the foot and ankle: diagnostic

- performance and patient acceptance of a dedicated low field MR scanner. J Magn Reson Imaging 8:711-6, .
- **149.** Kulkarni MM, Patton JJ, Price RR. Technical considerations for the use of surface coils in MRI. AJR Am J Roentgenol 147:373-8, .
- **150.** Rubin DA. MR imaging technique and principles. In: Spounge AR, Pope TL, ed. Practical MRI of the Foot and Ankle. Boca Raton, Fla: CRC Press LLC; 2001:1-36.
- **151.** Muhle C, Frank LR, Rand T, et al. Collateral ligaments of the ankle: high-resolution MR imaging with a local gradient coil and anatomic correlation in cadavers. Radiographics. 1999;19(3):673-83.
- **152.** Rubin DA, Kneeland JB. MR imaging of the musculoskeletal system: technical considerations for enhancing image quality and diagnostic yield. AJR Am J Roentgenol. 1994 Nov;163(5):1155-63.
- **153.** Feighan J, Towers J, Conti S. The use of magnetic resonance imaging in posterior tibial tendon dysfunction. Clin Orthop Relat Res. 1999 Aug;(365):23-38.
- **154.** Mengiardi BB, Pfirrmann CC, Schöttle PP, et al. Magic angle effect in MR imaging of ankle tendons: influence of foot positioning on prevalence and site in asymptomatic subjects and cadaveric tendons. Eur Radiol 16:2197-206, .
- **155.** Tokuda OO, Awaya HH, Taguchi KK, Matsunga NN. Kinematic MRI of the normal ankle ligaments using a specially designed passive positioning device. Foot Ankle Int 27:935-42, .
- 156. Hricak HH, Amparo EE. Body MRI: alleviation of claustrophobia by prone positioning. Radiology 152:819, .
- **157.** Rubin DD, Towers JJ, Britton CC. MR imaging of the foot: utility of complex oblique imaging planes. AJR Am J Roentgenol 166:1079-84, .
- **158.** Klein MM. Reformatted three-dimensional Fourier transform gradient-recalled echo MR imaging of the ankle: spectrum of normal and abnormal findings. AJR Am J Roentgenol 161:831-6, .
- **159.** Duc SS, Mengiardi BB, Pfirrmann CC, Hodler JJ, Zanetti MM. Improved visualization of collateral ligaments of the ankle: multiplanar reconstructions based on standard 2D turbo spin-echo MR images. Eur Radiol 17:1162-71, .
- **160.** Soila K, Karjalainen PT, Aronen HJ, Pihlajamäki HK, Tirman PJ. High-resolution MR imaging of the asymptomatic Achilles tendon: new observations. AJR Am J Roentgenol. 1999 Aug;173(2):323-8.
- **161.** Bush CH. The magnetic resonance imaging of musculoskeletal hemorrhage. Skeletal Radiol. 2000 Jan;29(1):1-9
- **162.** Ba-Ssalamah AA, Schibany NN, Puig SS, Herneth AA, Noebauer-Huhmann II, Trattnig SS. Imaging articular cartilage defects in the ankle joint with 3D fat-suppressed echo planar imaging: comparison with conventional 3D fat-suppressed gradient echo imaging. J Magn Reson Imaging 16:209-16, .
- **163.** Tan TT, Wilcox DD, Frank LL, et al. MR imaging of articular cartilage in the ankle: comparison of available imaging sequences and methods of measurement in cadavers. Skeletal Radiol 25:749-55, .
- **164.** Murphey MM, Rhee JJ, Lewis RR, Fanburg-Smith JJ, Flemming DD, Walker EE. Pigmented villonodular synovitis: radiologic-pathologic correlation. Radiographics 28:1493-518, .
- **165.** Jacobson JJ, Andresen RR, Jaovisidha SS, et al. Detection of ankle effusions: comparison study in cadavers using radiography, sonography, and MR imaging. AJR Am J Roentgenol 170:1231-8, .
- **166.** Huh YM, Suh JS, Lee JW, Song HT. Synovitis and soft tissue impingement of the ankle: assessment with enhanced three-dimensional FSPGR MR imaging. J Magn Reson Imaging. 2004;19(1):108-116.
- **167.** Kirchgesner T, Perlepe V, Michoux N, Larbi A, Vande Berg B. Fat suppression at 2D MR imaging of the hands: Dixon method versus CHESS technique and STIR sequence. Eur J Radiol. 2017 Apr;89():S0720-048X(17)30015-3.
- **168.** Fleckenstein JJ, Archer BB, Barker BB, Vaughan JJ, Parkey RR, Peshock RR. Fast short-tau inversion-recovery MR imaging. Radiology 179:499-504, .
- **169.** Fuller SS, Reeder SS, Shimakawa AA, et al. Iterative decomposition of water and fat with echo asymmetry and least-squares estimation (IDEAL) fast spin-echo imaging of the ankle: initial clinical experience. AJR Am J Roentgenol 187:1442-7, .
- **170.** Lu AA, Grist TT, Block WW. Fat/water separation in single acquisition steady-state free precession using multiple echo radial trajectories. Magn Reson Med 54:1051-7, .
- **171.** Maas MM, Dijkstra PP, Akkerman EE. Uniform fat suppression in hands and feet through the use of two-point Dixon chemical shift MR imaging. Radiology 210:189-93, .
- **172.** Reeder SS, McKenzie CC, Pineda AA, et al. Water-fat separation with IDEAL gradient-echo imaging. J Magn Reson Imaging 25:644-52, .
- **173.** Huo DD, Li ZZ, Aboussouan EE, Karis JJ, Pipe JJ. Turboprop IDEAL: a motion-resistant fat-water separation technique. Magn Reson Med 61:188-95, .
- **174.** Simon JJ, Szumowski JJ. Chemical shift imaging with paramagnetic contrast material enhancement for improved lesion depiction. Radiology 171:539-43, .
- 175. Bauer JJ, Banerjee SS, Henning TT, Krug RR, Majumdar SS, Link TT. Fast high-spatial-resolution MRI of the

ankle with parallel imaging using GRAPPA at 3 T. AJR Am J Roentgenol 189:240-5, .

R**१९**। See २०२३ KR: Bolsseं परार्भी an EE, et al. Ankle: isotropic MR imaging with 3D-FSE-cube--initial experience in healthy volunteers. Radiology 249:1026-33, .

- **177.** Peh WC, Chan JH. Artifacts in musculoskeletal magnetic resonance imaging: identification and correction. Skeletal Radiol. 2001 Apr;30(4):179-91.
- **178.** Van Hecke PE, Marchal GJ, Baert AL. Use of shielding to prevent folding in MR imaging. Radiology. 1988 May;167(2):557-8.
- **179.** Haacke EM, Lenz GW. Improving MR image quality in the presence of motion by using rephasing gradients. AJR Am J Roentgenol. 1987 Jun;148(6):1251-8.
- **180.** Du JJ, Pak BB, Znamirowski RR, et al. Magic angle effect in magnetic resonance imaging of the Achilles tendon and enthesis. Magn Reson Imaging 27:557-64, .
- **181.** Erickson SS, Cox II, Hyde JJ, Carrera GG, Strandt JJ, Estkowski LL. Effect of tendon orientation on MR imaging signal intensity: a manifestation of the "magic angle" phenomenon. Radiology 181:389-92, .
- **182.** Peh WW, Chan JJ. The magic angle phenomenon in tendons: effect of varying the MR echo time. Br J Radiol 71:31-6, .
- **183.** Schulte-Altedorneburg G, Gebhard M, Wohlgemuth WA, et al. MR arthrography: pharmacology, efficacy and safety in clinical trials. Skeletal Radiol. 2003 Jan;32(1):1-12.
- **184.** American College of Radiology. ACR Practice Parameter for Communication of Diagnostic Imaging Findings. Available at https://gravitas.acr.org/PPTS/GetDocumentView?docId=74+&releaseId=2
- **185.** Shellock FG. Guide to MR Procedures and Metallic Objects: Update 2001. 7th ed. Philadelphia, Pa: Lippincott Williams & Wilkins; 2001.
- **186.** Shellock FG, Spinazzi A. MRI safety update 2008: part 2, screening patients for MRI. AJR Am J Roentgenol. 2008 Oct;191(4):1140-9.
- **187.** American College of Radiology. ACR–AAPM Technical Standard for Diagnostic Medical Physics Performance Monitoring of Magnetic Resonance (MR) Imaging Equipment. Available at https://gravitas.acr.org/PPTS/GetDocumentView?docId=57+&releaseId=2
- *Practice parameters and technical standards are published annually with an effective date of October 1 in the year in which amended, revised or approved by the ACR Council. For practice parameters and technical standards published before 1999, the effective date was January 1 following the year in which the practice parameter or technical standard was amended, revised, or approved by the ACR Council.

<u>Development Chronology for this Practice Parameter</u>

2006 (Resolution 4, 35)

Revised 2011 (Resolution 20)

Amended 2014 (Resolution 39)

Revised 2016 (Resolution 3)

Revised 2021 (Resolution 44)

Amended 2023 (Resolution 2c)