

ACR–AAPM TECHNICAL STANDARD FOR NUCLEAR MEDICAL PHYSICS PERFORMANCE MONITORING OF GAMMA CAMERAS

The American College of Radiology, with more than 40,000 members, is the principal organization of radiologists, radiation oncologists, and clinical medical physicists in the United States. The College is a nonprofit professional society whose primary purposes are to advance the science of radiology, improve radiologic services to the patient, study the socioeconomic aspects of the practice of radiology, and encourage continuing education for radiologists, radiation oncologists, medical physicists, and persons practicing in allied professional fields.

The American College of Radiology will periodically define new practice parameters and technical standards for radiologic practice to help advance the science of radiology and to improve the quality of service to patients throughout the United States. Existing practice parameters and technical standards will be reviewed for revision or renewal, as appropriate, on their fifth anniversary or sooner, if indicated.

Each practice parameter and technical standard, representing a policy statement by the College, has undergone a thorough consensus process in which it has been subjected to extensive review and approval. The practice parameters and technical standards recognize that the safe and effective use of diagnostic and therapeutic radiology requires specific training, skills, and techniques, as described in each document. Reproduction or modification of the published practice parameter and technical standard by those entities not providing these services is not authorized.

PREAMBLE

This document is an educational tool designed to assist practitioners in providing appropriate radiologic care for patients. Practice Parameters and Technical Standards are not inflexible rules or requirements of practice and are not intended, nor should they be used, to establish a legal standard of care¹. For these reasons and those set forth below, the American College of Radiology and our collaborating medical specialty societies caution against the use of these documents in litigation in which the clinical decisions of a practitioner are called into question.

The ultimate judgment regarding the propriety of any specific procedure or course of action must be made by the practitioner considering all the circumstances presented. Thus, an approach that differs from the guidance in this document, standing alone, does not necessarily imply that the approach was below the standard of care. To the contrary, a conscientious practitioner may responsibly adopt a course of action different from that set forth in this document when, in the reasonable judgment of the practitioner, such course of action is indicated by variables such as the condition of the patient, limitations of available resources, or advances in knowledge or technology after publication of this document. However, a practitioner who employs an approach substantially different from the guidance in this document may consider documenting in the patient record information sufficient to explain the approach taken.

The practice of medicine involves the science, and the art of dealing with the prevention, diagnosis, alleviation, and treatment of disease. The variety and complexity of human conditions make it impossible to always reach the most appropriate diagnosis or to predict with certainty a particular response to treatment. Therefore, it should be recognized that adherence to the guidance in this document will not assure an accurate diagnosis or a successful outcome. All that should be expected is that the practitioner will follow a reasonable course of action based on current knowledge, available resources, and the needs of the patient to deliver effective and safe medical care. The purpose of this document is to assist practitioners in achieving this objective.

¹ *Iowa Medical Society and Iowa Society of Anesthesiologists v. Iowa Board of Nursing*, 831 N.W.2d 826 (Iowa 2013) Iowa Supreme Court refuses to find that the "ACR Technical Standard for Management of the Use of Radiation in Fluoroscopic Procedures (Revised 2008)" sets a national standard for who may perform fluoroscopic procedures in light of the standard's stated purpose that ACR standards are educational tools and not intended to establish a legal standard of care. See also, *Stanley v. McCarver*, 63 P.3d 1076 (Ariz. App. 2003) where in a concurring opinion the Court stated that "published standards or guidelines of specialty medical organizations are useful in determining the duty owed or the standard of care applicable in a given situation" even though ACR standards themselves do not establish the standard of care.

I. INTRODUCTION

This technical standard was revised collaboratively by the American College of Radiology (ACR) and the American Association of Physicists in Medicine (AAPM).

All nuclear medicine imaging equipment must be tested upon installation and monitored at least annually by a Qualified Medical Physicist to ensure that it is functioning within manufacturer specifications and accepted performance standards. Additional or more frequent performance monitoring may be necessary in certain situations (eg, after major equipment maintenance). Although it is not possible to consider all variations of equipment performance to be monitored, adherence to this technical standard will optimize image quality and help to improve the accuracy of quantitative results in clinical procedures. Key points to consider are performance characteristics to be monitored, estimated patient radiation dose, qualifications of personnel, and follow-up procedures.

II. QUALIFICATIONS AND RESPONSIBILITIES OF PERSONNEL

A Qualified Medical Physicist must carry out acceptance testing and monitoring of gamma camera equipment.

A Qualified Medical Physicist is an individual who is competent to practice independently one or more of the subfields in medical physics. The American College of Radiology (ACR) considers certification, continuing education and experience in the appropriate subfield(s) to demonstrate that an individual is competent to practice one or more of the subfields in medical physics and to be a Qualified Medical Physicist. The ACR strongly recommends that the individual be certified in the appropriate subfield(s) by the American Board of Radiology (ABR), the Canadian College of Physicists in Medicine, the American Board of Science in Nuclear Medicine (ABSNM), or the American Board of Medical Physics (ABMP).

A Qualified Medical Physicist should meet the [ACR Practice Parameter for Continuing Medical Education \(CME\)](#) [1].

The appropriate subfield of medical physics for this standard is Nuclear Medical Physics (including medical physics certification categories of Radiological Physics, Medical Nuclear Physics, and Nuclear Medicine Physics).

(ACR Resolution 17, adopted in 1996 – revised in 2008, 2012, 2022, Resolution 41f)

The Qualified Medical Physicist must be familiar with the principles of radiation protection; the guidelines of the National Council on Radiation Protection and Measurements (NCRP); laws and regulations governing the use of the equipment being tested; the function, clinical uses, and performance specifications of the imaging equipment; and calibration processes and limitations of the instruments and techniques used for testing performance.

The Qualified Medical Physicist is responsible for:

1. The design of the overall program of performance monitoring (including the selection of specific methods for acceptance testing and quality control testing)
2. Documentation of program goals, policies, and procedures related to performance monitoring
3. Documentation of the results of all performance measurements
4. Review and approval of all measurements performed by other designated personnel

The Qualified Medical Physicist may be assisted by properly trained individuals in obtaining data for performance monitoring. These individuals must be approved by the Qualified Medical Physicist in the techniques of performing tests.

III. PERFORMANCE CHARACTERISTICS TO BE MONITORED

A. Acceptance Testing

Initial performance testing of imaging equipment must be performed upon installation and should be completed before clinical use. This testing should be more comprehensive than periodic performance testing and should be consistent with current acceptance testing practices [2]. Electrical safety of the equipment must also be tested by appropriate personnel prior to its initial clinical use.

B. Performance Evaluation

Gamma Camera may be a single head, dual-head, or multi-head system. Gamma Camera systems may also be of a continuous crystal or "pixelated" design. The following characteristics should be evaluated at least annually [3-16]. The evaluation of some of these characteristics may not be applicable depending on the system design. In such cases, these evaluations may be removed or modified at the discretion of the Qualified Medical Physicist in accordance with manufacturer recommendations.

1. Intrinsic uniformity (for each gamma camera head)
2. Extrinsic uniformity with all collimators (for each gamma camera head)
3. Intrinsic or extrinsic spatial resolution
4. Intrinsic or extrinsic geometric distortion (also known as spatial linearity)
5. Extrinsic sensitivity
 - a. Count rate per unit activity
 - b. Interdetector variability
6. Energy Resolution
7. Count rate performance
8. Monitors used to display acquired images [17,18]
9. Overall system performance for single photon emission computed tomography (SPECT) if applicable [19].
 - a. Uniformity and artifact evaluation
 - b. Cold sphere contrast
 - c. Spatial resolution
10. Physical inspection, system interlocks and emergency shutdown mechanisms
11. Review of the gamma camera quality control program

C. Quality Management Program

A quality management (QM) program must be established for the nuclear medicine imaging equipment with the assistance of a Qualified Medical Physicist as outlined in the [ACR-ACNM-SNMMI-SPR Practice Parameter for the Use of Radiopharmaceuticals in Diagnostic Procedures](#) [20]. An on-site technologist should be identified to be responsible for conducting routine quality control (QC).

The results of the QM program must be monitored at least annually by the Qualified Medical Physicist. If measured values of QC/quality assurance (QA) parameters fall outside the established tolerances, the QC/QA technologist should consult with the Qualified Medical Physicist. The Qualified Medical Physicist should recommend or, when appropriate, initiate investigative or corrective actions. A Qualified Medical Physicist should be available to assist in prescribing corrective actions for unresolved problems.

D. Written Survey Reports and Follow-Up Procedures

The Qualified Medical Physicist must provide a written report of the findings of acceptance testing and a performance evaluation to the professional(s) in charge of obtaining or providing necessary service to the equipment and, if appropriate, to the responsible physician(s). Written reports must be provided in a timely manner consistent with the importance of any adverse findings.

If appropriate, the Qualified Medical Physicist should notify the facility to initiate the required service. The facility must complete corrective actions in a timely manner consistent with the importance of any adverse findings. The facility should retain service reports from competent service personnel as verification that the issue(s) were appropriately resolved. The reports may be reviewed by a Qualified Medical Physicist to confirm that the equipment is performing in a safe and acceptable fashion after the required service is performed or as required by federal, state, or local regulations.

If use of the equipment would pose a danger to life or health or potentially result in erroneous clinical findings, the Qualified Medical Physicist in collaboration with the facility's Radiation Safety Officer and

interpreting physician must take immediate action to either prevent equipment use or to indicate in writing what limited studies can be performed safely using the equipment until the hazard is addressed.

IV. RADIATION SAFETY IN IMAGING

Radiologists, medical physicists, non-physician radiology providers, radiologic technologists, and all supervising physicians have a responsibility for safety in the workplace by keeping radiation exposure to staff, and to society as a whole, "as low as reasonably achievable" (ALARA) and to assure that radiation doses to individual patients are appropriate, taking into account the possible risk from radiation exposure and the diagnostic image quality necessary to achieve the clinical objective. All personnel who work with ionizing radiation must understand the key principles of occupational and public radiation protection (justification, optimization of protection, application of dose constraints and limits) and the principles of proper management of radiation dose to patients (justification, optimization including the use of dose reference levels). https://www-pub.iaea.org/MTCD/Publications/PDF/PUB1775_web.pdf

Facilities and their responsible staff should consult with the radiation safety officer to ensure that there are policies and procedures for the safe handling and administration of radiopharmaceuticals in accordance with ALARA principles. These policies and procedures must comply with all applicable radiation safety regulations and conditions of licensure imposed by the Nuclear Regulatory Commission (NRC) and by applicable state, local, or other relevant regulatory agencies and accrediting bodies, as appropriate. Quantities of radiopharmaceuticals should be tailored to the individual patient by prescription or protocol, using body habitus or other customized method when such guidance is available.

Nationally developed guidelines, such as the [ACR's Appropriateness Criteria](#)[®], should be used to help choose the most appropriate imaging procedures to prevent unnecessary radiation exposure.

Additional information regarding patient radiation safety in imaging is available from the following websites – Image Gently[®] for children (www.imagegently.org) and Image Wisely[®] for adults (www.imagewisely.org). These advocacy and awareness campaigns provide free educational materials for all stakeholders involved in imaging (patients, technologists, referring providers, medical physicists, and radiologists).

Radiation exposures or other dose indices should be periodically measured by a Qualified Medical Physicist in accordance with the applicable ACR Technical Standards. Monitoring or regular review of dose indices from patient imaging should be performed by comparing the facility's dose information with national benchmarks, such as the ACR Dose Index Registry and relevant publications relying on its data, applicable ACR Practice Parameters, NCRP Report No. 172, Reference Levels and Achievable Doses in Medical and Dental Imaging: Recommendations for the United States or the Conference of Radiation Control Program Director's National Evaluation of X-ray Trends; 2006, 2009, amended 2013, revised 2023 (Res. 2d).

Information on typical organ doses from radiopharmaceuticals should be available for all procedures. This information must be reviewed by the Qualified Medical Physicist at least annually and updated when any of the following occur: addition of new procedures and/or pharmaceuticals, changes in dosage schedules, change in route of administration, and availability of more accurate dosimetry data [21-25]. For facilities performing pediatric imaging, the radiopharmaceutical administered activities should be adjusted to be appropriate for the mass of the patient. It is recommended that the administered activities follow those in the North American Consensus Guidelines for Administered Radiopharmaceutical Activities in Children and Adolescents [15,16,26,27].

ACKNOWLEDGEMENTS

This technical standard was revised according to the process described under the heading *the Process for Developing ACR Practice Guidelines and Technical Standards* on the ACR website (<https://www.acr.org/Clinical-Resources/Practice-Parameters-and-Technical-Standards>) by the Committee on Practice Parameters and Technical Standards – Medical Physics of the ACR Commission on Medical Physics in collaboration with the AAPM.

Writing Committee – members represent their societies in the initial and final revision of this technical standard

Ashley E. Rubinstein, PhD, Chair

Carrie B. Hruska, PhD

Maxwell R. Amurao, PhD, MBA

Osama Mawlawi, PhD, FACR

Tariq A. Mian, PhD, FACR

Pankaj Patel, PhD

Committee on Practice Parameters and Technical Standards – Medical Physics

(ACR Committee responsible for sponsoring the draft through the process)

Mary Ann Keenan, DMP, Chair

Samuel A. Einstein, PhD

Katherine P Andriole, PhD, FACR

Ralph P. Lieto, MSE, FACR

Eric A. Berns, Ph.D, FACR

Osama Mawlawi, PhD, FACR

Courtney R. Buckey, PhD

Matthew A. Pacella, MS, FACR

Diana E. Carver, PhD

Ashley E. Rubinstein, PhD

Heidi A. Edmonson, PhD

Russell B. Tarver, MS

Mahadevappa Mahesh, MS, PhD, FACR, Chair, Commission on Medical Physics

David B. Larson, MD, MBA, FACR, Chair, Commission on Quality and Safety

Mary S. Newell, MD, FACR, Chair, Committee on Practice Parameters and Technical Standards

Comments Reconciliation Committee

Maxwell R. Amurao, PhD, MBA, Chair

Mahadevappa Mahesh, MS, PhD, FACR

K. Elizabeth Hawk, MD, MS, PhD, Co-Chair

Gustavo A. Mercier, Jr., MD, PhD

Timothy A. Crummy, MD, MHA, FACR

Osama Mawlawi, PhD, FACR

Comments Reconciliation Committee

William Erwin, MS

Tariq A. Mian, PhD, FACR

Carrie B. Hruska, PhD

Mary S. Newell, MD, FACR

Mary Ellen Jafari, MD, FACR

Pankaj Patel, PhD

Mary Ann Keenan, DMP

Ashley E. Rubinstein, PhD

Amy L. Kotsenas, MD, FACR

Roland Wong, ScM

David B. Larson, MD, MBA, FACR

REFERENCES

1. American College of Radiology. ACR practice parameter for continuing medical education (CME). Available at: <https://www.acr.org/-/media/ACR/Files/Practice-Parameters/CME.pdf>. Accessed January 14, 2022.
2. American Association of Physicists in Medicine. The Report of AAPM Task Group 177: acceptance testing and annual physics survey recommendations for gamma camera, SPECT, and SPECT/CT systems (2019). Available at: https://www.aapm.org/pubs/reports/RPT_177.pdf. Accessed June 6, 2022.
3. *Quality Assurance for Radioactivity Measurement in Nuclear Medicine*. Vienna, Austria: International Atomic Energy Agency; 2006. Technical Reports Series No. 454.
4. *Performance Measurements of Gamma Cameras*. District of Columbia, DC: National Electrical Manufacturers Association; 2007. Standards Publication NU 1-2007.
5. Bolster A. *Quality Assurance in Gamma Camera Systems*: Institute of Physics and Engineering in Medicine; 2003. Report 86.
6. Forstrom LA, Dunn WL, O'Connor MK, Decklever TD, Hardyman TJ, Howarth DM. Technical pitfalls in image acquisition, processing, and display. *Seminars in nuclear medicine* 1996;26:278-94.
7. Graham LS, Fahey FH, Madsen MT, van Aswegen A, Yester MV. Quantitation of SPECT performance: Report of Task Group 4, Nuclear Medicine Committee. *Medical physics* 1995;22:401-9.
8. Henkin R ed. *Nuclear Principles and Practices*. St. Louis, Mo: Mosby-Yearbook; 1996.
9. Hines H, Kayayan R, Colsher J, et al. National Electrical Manufacturers Association recommendations for implementing SPECT instrumentation quality control. *Journal of nuclear medicine : official publication, Society of Nuclear Medicine* 2000;41:383-9.
10. International Atomic Energy Agency. IAEA Quality Control Atlas for Scintillation Camera Systems. http://www-pub.iaea.org/MTCD/publications/PDF/Pub1141_web.pdf. Accessed January 30, 2012.
11. O'Connor MK. Instrument- and computer-related problems and artifacts in nuclear medicine. *Seminars in nuclear medicine* 1996;26:256-77.
12. Sandler M ed. *Diagnostic Nuclear Medicine*. 4th ed. Baltimore, Md: Lippincott, Williams and Wilkins; 2003.
13. Siman W, Kappadath SC. Performance characteristics of a new pixelated portable gamma camera. *Medical physics* 2012;39:3435-44.
14. Garcia EV, Faber TL. New trends in camera and software technology in nuclear cardiology. *Cardiology clinics* 2009;27:227-36, Table of Contents.
15. Lassmann M, Treves ST. Pediatric Radiopharmaceutical Administration: harmonization of the 2007 EANM Paediatric Dosage Card (Version 1.5.2008) and the 2010 North American Consensus guideline. *Eur J Nucl Med Mol Imaging* 2014;41:1636.

16. Treves ST, Gelfand MJ, Fahey FH, Parisi MT. 2016 Update of the North American Consensus Guidelines for Pediatric Administered Radiopharmaceutical Activities. Journal of nuclear medicine : official publication, Society of Nuclear Medicine 2016;57:15N-18N.
17. American Association of Physicists in Medicine. The Report of AAPM Task Group 270: display quality assurance (2019). Available at: https://www.aapm.org/pubs/reports/RPT_270.pdf. Accessed April 22, 2022.
18. American College of Radiology. ACR–AAPM–SIIM technical standard for electronic practice of medical imaging. Available at: <https://www.acr.org/-/media/ACR/Files/Practice-Parameters/Elec-Practice-MedImag.pdf>. Accessed April 22, 2022.
19. American College of Radiology. ACR–ACNM–SNMMI–SPR practice parameter for the performance of single-photon emission brain perfusion imaging (including SPECT and SPECT/CT). Available at: <https://www.acr.org/-/media/ACR/Files/Practice-Parameters/BrainPerf-SPECT.pdf>. Accessed April 22, 2022.
20. American College of Radiology. ACR–ACNM–SNMMI–SPR practice parameter for the use of radiopharmaceuticals in diagnostic procedures. Available at: <https://www.acr.org/-/media/ACR/Files/Practice-Parameters/Radiopharmaceuticals.pdf>. Accessed January 14, 2022.
21. American College of Radiology. ACR–AAPM–ACNM–SNMMI practice parameter for reference levels and achievable administered activity for nuclear medicine and molecular imaging. Available at: <https://www.acr.org/-/media/ACR/Files/Practice-Parameters/RefLevels-NucMed.pdf>. Accessed October 26, 2022.
22. International Commission on Radiological Protection. Publication 53 - radiation dose to patients from radiopharmaceuticals. Available at: <https://www.icrp.org/publication.asp?id=ICRP%20Publication%2053>. Accessed October 26, 2022.
23. International Commission on Radiological Protection. Publication 80 - radiation dose to patients from radiopharmaceuticals addendum to ICRP 53; also includes addendum 1 to ICRP publication 72. Available at: https://journals.sagepub.com/doi/pdf/10.1177/ANIB_28_3. Accessed October 26, 2022.
24. International Commission on Radiological Protection. Publication 106 - radiation dose to patients from radiopharmaceuticals (a third amendment to ICRP Publication 53; also includes: radiation exposure of hands in radiopharmacies). Available at: https://journals.sagepub.com/doi/pdf/10.1177/ANIB_38_1-2. Accessed October 26, 2022.
25. International Commission on Radiological Protection. Publication 128 - radiation dose to patients from radiopharmaceuticals: a compendium of current information related to frequently used substances. Available at: https://journals.sagepub.com/doi/pdf/10.1177/ANIB_44_2S. Accessed October 26, 2022.
26. Oak Ridge Institute for Science and Education. Dose estimates and other compendia. Available at: <https://orise.ornl.gov/resources/reacts/dose-estimation-resources.html>. Accessed June 14, 2022.
27. Society of Nuclear Medicine and Molecular Imaging. Dosimetry. Available at: <http://www.snmmi.org/ClinicalPractice/content.aspx?ItemNumber=5325&navItemNumber=10792>. Accessed February 15, 2017.

*As of May 2015, all practice parameters and technical standards that are collaborative with only the American Association of Physics in Medicine are approved by the ACR Council Steering Committee and the ACR Board of Chancellors and will not go through the ACR Council (ACR Resolution 54, 2015). The effective date is the first day of the month following a 60-day period that begins on the date the document was approved.

Development Chronology for this Technical Standard 1998 (Resolution 16)

Revised 2003 (Resolution 14)

Amended 2006 (Resolution 16g, 17)

Revised 2008 (Resolution 5)

Amended 2012 (Resolution 42)

Revised 2013 (Resolution 42)

Revised 2018 (CSC/BOC)

Revised 2023 (CSC/BOC) (Revision 41f)

Revised 2023 (CSC/BOC)